

| PROTEIN | CALIBRATOR 1 | CALIBRATOR 2 | CALIBRATOR 3 | CALIBRATOR 4 | CALIBRATOR 5 |
|-------------------------------|--------------|--------------|--------------|--------------|--------------|
| ALBUMIN | 7.30 g/l | 13.80 g/l | 25.80 g/l | 50.90 g/l | 100.50 g/l |
| α -1 ANTITRYPSIN | 0.21 g/l | 0.41 g/l | 0.81 g/l | 1.62 g/l | 3.25 g/l |
| α -1 GLYCOPROTEIN ACID | 0.12 g/l | 0.23 g/l | 0.46 g/l | 0.92 g/l | 1.84 g/l |
| α -2 MACROGLOBULIN | 0.36 g/l | 0.68 g/l | 1.35 g/l | 2.72 g/l | 5.43 g/l |
| ANTITHROMBIN III | 0.037 g/l | 0.084 g/l | 0.191 g/l | 0.416 g/l | 0.867 g/l |
| COMPLEMENT C3 | 0.23 g/l | 0.45 g/l | 0.89 g/l | 1.79 g/l | 3.58 g/l |
| COMPLEMENT C4 | 0.042 g/l | 0.083 g/l | 0.163 g/l | 0.327 g/l | 0.656 g/l |
| CERULOPLASMIN* | 0.054 g/l | 0.105 g/l | 0.214 g/l | 0.464 g/l | 0.850 g/l |
| CERULOPLASMIN** | 0.059 g/l | 0.119 g/l | 0.239 g/l | 0.505 g/l | 1.081 g/l |
| CERULOPLASMIN*** | 0.093 g/l | 0.158 g/l | 0.304 g/l | 0.621 g/l | 1.214 g/l |
| HAPTOGLOBIN | 0.18 g/l | 0.36 g/l | 0.72 g/l | 1.45 g/l | 2.90 g/l |
| IgA | 0.31 g/l | 0.62 g/l | 1.23 g/l | 2.46 g/l | 4.94 g/l |
| IgG | 1.61 g/l | 3.18 g/l | 6.30 g/l | 12.61 g/l | 25.28 g/l |
| IgM | 0.15 g/l | 0.31 g/l | 0.60 g/l | 1.21 g/l | 2.43 g/l |
| PREALBUMIN | 0.037 g/l | 0.074 g/l | 0.157 g/l | 0.324 g/l | 0.682 g/l |
| TRANSFERRIN | 0.41 g/l | 0.81 g/l | 1.61 g/l | 3.21 g/l | 6.44 g/l |

*Values compatible with CERULOPLASMIN reagents LOT number xxx-xx23 and xxx-xx24

**Values compatible with CERULOPLASMIN reagents LOT number xxx-xx25 and xxx-xx26

***Values compatible with CERULOPLASMIN reagents since LOT number xxx-xx27

| PROTEIN | CALIBRATOR 1 | CALIBRATOR 2 | CALIBRATOR 3 | CALIBRATOR 4 | CALIBRATOR 5 |
|-------------------------------|--------------|--------------|--------------|--------------|--------------|
| ALBUMIN | 7.30 g/l | 13.80 g/l | 25.80 g/l | 50.90 g/l | 100.50 g/l |
| α -1 ANTITRYPSIN | 0.21 g/l | 0.41 g/l | 0.81 g/l | 1.62 g/l | 3.25 g/l |
| α -1 GLYCOPROTEIN ACID | 0.12 g/l | 0.23 g/l | 0.46 g/l | 0.92 g/l | 1.84 g/l |
| α -2 MACROGLOBULIN | 0.36 g/l | 0.68 g/l | 1.35 g/l | 2.72 g/l | 5.43 g/l |
| ANTITHROMBIN III | 0.037 g/l | 0.084 g/l | 0.191 g/l | 0.416 g/l | 0.867 g/l |
| COMPLEMENT C3 | 0.23 g/l | 0.45 g/l | 0.89 g/l | 1.79 g/l | 3.58 g/l |
| COMPLEMENT C4 | 0.042 g/l | 0.083 g/l | 0.163 g/l | 0.327 g/l | 0.656 g/l |
| CERULOPLASMIN* | 0.054 g/l | 0.105 g/l | 0.214 g/l | 0.464 g/l | 0.850 g/l |
| CERULOPLASMIN** | 0.059 g/l | 0.119 g/l | 0.239 g/l | 0.505 g/l | 1.081 g/l |
| CERULOPLASMIN*** | 0.093 g/l | 0.158 g/l | 0.304 g/l | 0.621 g/l | 1.214 g/l |
| HAPTOGLOBIN | 0.18 g/l | 0.36 g/l | 0.72 g/l | 1.45 g/l | 2.90 g/l |
| IgA | 0.31 g/l | 0.62 g/l | 1.23 g/l | 2.46 g/l | 4.94 g/l |
| IgG | 1.61 g/l | 3.18 g/l | 6.30 g/l | 12.61 g/l | 25.28 g/l |
| IgM | 0.15 g/l | 0.31 g/l | 0.60 g/l | 1.21 g/l | 2.43 g/l |
| PREALBUMIN | 0.037 g/l | 0.074 g/l | 0.157 g/l | 0.324 g/l | 0.682 g/l |
| TRANSFERRIN | 0.41 g/l | 0.81 g/l | 1.61 g/l | 3.21 g/l | 6.44 g/l |

*Values compatible with CERULOPLASMIN reagents LOT number xxx-xx23 and xxx-xx24

**Values compatible with CERULOPLASMIN reagents LOT number xxx-xx25 and xxx-xx26

***Values compatible with CERULOPLASMIN reagents since LOT number xxx-xx27

| BIALKO | CALIBRATOR 1 | CALIBRATOR 2 | CALIBRATOR 3 | CALIBRATOR 4 | CALIBRATOR 5 |
|---------------------------------|--------------|--------------|--------------|--------------|--------------|
| ALBUMINA | 7,30 g/l | 13,80 g/l | 25,80 g/l | 50,90 g/l | 100,50 g/l |
| α -1 ANTYTRYPSYNA | 0,21 g/l | 0,41 g/l | 0,81 g/l | 1,62 g/l | 3,25 g/l |
| KWAŚNA α 1-GLIKOPROTEINA | 0,12 g/l | 0,23 g/l | 0,46 g/l | 0,92 g/l | 1,84 g/l |
| α -2 MAKROGLOBULINA | 0,36 g/l | 0,68 g/l | 1,35 g/l | 2,72 g/l | 5,43 g/l |
| ANTYTROMBINA III | 0,037 g/l | 0,084 g/l | 0,191 g/l | 0,416 g/l | 0,867 g/l |
| SKŁADNIK C3 DOPEŁNIACZA | 0,23 g/l | 0,45 g/l | 0,89 g/l | 1,79 g/l | 3,58 g/l |
| SKŁADNIK C4 DOPEŁNIACZA | 0,042 g/l | 0,083 g/l | 0,163 g/l | 0,327 g/l | 0,656 g/l |
| CERULOPLAZMINA* | 0,054 g/l | 0,105 g/l | 0,214 g/l | 0,464 g/l | 0,850 g/l |
| CERULOPLAZMINA** | 0,059 g/l | 0,119 g/l | 0,239 g/l | 0,505 g/l | 1,081 g/l |
| CERULOPLAZMINA*** | 0,093 g/l | 0,158 g/l | 0,304 g/l | 0,621 g/l | 1,214 g/l |
| HAPTOGLOBINA | 0,18 g/l | 0,36 g/l | 0,72 g/l | 1,45 g/l | 2,90 g/l |
| IgA | 0,31 g/l | 0,62 g/l | 1,23 g/l | 2,46 g/l | 4,94 g/l |
| IgG | 1,61 g/l | 3,18 g/l | 6,30 g/l | 12,61 g/l | 25,28 g/l |
| IgM | 0,15 g/l | 0,31 g/l | 0,60 g/l | 1,21 g/l | 2,43 g/l |
| PREALBUMINA | 0,037 g/l | 0,074 g/l | 0,157 g/l | 0,324 g/l | 0,682 g/l |
| TRANSFERYNA | 0,41 g/l | 0,81 g/l | 1,61 g/l | 3,21 g/l | 6,44 g/l |

*Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie xxx-xx23 oraz xxx-xx24

**Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie xxx-xx25 oraz xxx-xx26

***Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie od xxx-xx27

| BIALKO | CALIBRATOR 1 | CALIBRATOR 2 | CALIBRATOR 3 | CALIBRATOR 4 | CALIBRATOR 5 |
|---------------------------------|--------------|--------------|--------------|--------------|--------------|
| ALBUMINA | 7,30 g/l | 13,80 g/l | 25,80 g/l | 50,90 g/l | 100,50 g/l |
| α -1 ANTYTRYPSYNA | 0,21 g/l | 0,41 g/l | 0,81 g/l | 1,62 g/l | 3,25 g/l |
| KWAŚNA α 1-GLIKOPROTEINA | 0,12 g/l | 0,23 g/l | 0,46 g/l | 0,92 g/l | 1,84 g/l |
| α -2 MAKROGLOBULINA | 0,36 g/l | 0,68 g/l | 1,35 g/l | 2,72 g/l | 5,43 g/l |
| ANTYTROMBINA III | 0,037 g/l | 0,084 g/l | 0,191 g/l | 0,416 g/l | 0,867 g/l |
| SKŁADNIK C3 DOPEŁNIACZA | 0,23 g/l | 0,45 g/l | 0,89 g/l | 1,79 g/l | 3,58 g/l |
| SKŁADNIK C4 DOPEŁNIACZA | 0,042 g/l | 0,083 g/l | 0,163 g/l | 0,327 g/l | 0,656 g/l |
| CERULOPLAZMINA* | 0,054 g/l | 0,105 g/l | 0,214 g/l | 0,464 g/l | 0,850 g/l |
| CERULOPLAZMINA** | 0,059 g/l | 0,119 g/l | 0,239 g/l | 0,505 g/l | 1,081 g/l |
| CERULOPLAZMINA*** | 0,093 g/l | 0,158 g/l | 0,304 g/l | 0,621 g/l | 1,214 g/l |
| HAPTOGLOBINA | 0,18 g/l | 0,36 g/l | 0,72 g/l | 1,45 g/l | 2,90 g/l |
| IgA | 0,31 g/l | 0,62 g/l | 1,23 g/l | 2,46 g/l | 4,94 g/l |
| IgG | 1,61 g/l | 3,18 g/l | 6,30 g/l | 12,61 g/l | 25,28 g/l |
| IgM | 0,15 g/l | 0,31 g/l | 0,60 g/l | 1,21 g/l | 2,43 g/l |
| PREALBUMINA | 0,037 g/l | 0,074 g/l | 0,157 g/l | 0,324 g/l | 0,682 g/l |
| TRANSFERYNA | 0,41 g/l | 0,81 g/l | 1,61 g/l | 3,21 g/l | 6,44 g/l |

*Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie xxx-xx23 oraz xxx-xx24

**Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie xxx-xx25 oraz xxx-xx26

***Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie od xxx-xx27